


COMPLETE GAS SORPTION ANALYZER WITH MASS SPECTROMETER

Autosorb®-1-C/MS Benefits

- Combines gas detection by mass spectrometry with automatic gas sorption analysis to provide a more complete surface characterization of porous solids.
- Provides In-depth catalyst characterization, including micropore size determination in a single instrument.
- Achieves detailed identification of desorbed species under controlled heating rates and atmospheres.
- Requires little bench space because of integrated design utilizing Autosorb's proven vacuum system.

AUTOSORB®-1-C/MS



The exposed surface of many powders is often found deep within individual particles. This fact has a tremendous impact on applications of powdered solids because, even for a single solid, not all surfaces are created equal. Real surfaces are heterogeneous and their properties vary from exterior to interior as well as from one surface patch to another. Such variations can be more fully quantified by means of a complete gas sorption analyzer such as the **Autosorb®-1-C/MS**.

Temperature Programming

During a chemisorption analysis, the total number of active sites, or the active surface area, is evaluated. If one heats up the sample, the chemisorbed species tend to desorb according to how strongly they were bound to the active sites to begin with. Weaker sites normally release their chemisorbed species at lower temperatures than stronger sites.

If the sample had two distinct types of active sites, heating would release gases at two different temperatures. To quantify these releases, one can increase the sample temperature linearly and monitor the evolution of gases in a number of ways. If only a single gas is evolved, one can simply monitor the rise in pres-

sure in a closed, pre-evacuated sample cell as a function of temperature. This could be regarded as a "Soak" Temperature Programmed Desorption (TPD) experiment.

If more than one species was desorbed, it would be necessary to identify each one as it is evolved from the sample. This can be done by attaching a Mass Spectrometer (MS) to the exit valve of the **Autosorb®-1-C** and performing the thermal desorption under a flow of inert or carrier gas. As such, this experiment would be termed Flow TPD, or Flow TPR if a reactive gas is used instead of an inert gas. With MS sampling rates in the order of seconds, this allows the evaluation in detail of fast-releasing "reactive" sites, or readsorption phenomena.

Moreover, a properly placed MS probe can sample gases from an evacuated chamber (Vacuum TPD). A clear advantage of Vacuum TPD is that desorbing species can reach the MS more quickly and without being affected by diffusion or readsorption effects through pore networks.

Over...

The AUTOSORB-1-C/MS

The Autosorb[®]-1-C/MS provides physisorption, chemisorption, flow or vacuum TPD and residual gas analysis (MS) capabilities in a single unit.

1. Temperature programmed furnace: 1°C steps up to 1000°C with ramp rates up to 50°C per minute.
2. Quartz flow-through cell for chemisorption studies.
3. Floating cell guard.
4. Flexible positioning of quadrupole sensor and electronics control unit for various models of residual gas analyzers.
5. Needle valve for mass spectrometer or optional diversion valve for GC port.



WORLDWIDE Sales and Service

Argentina
Australia
Austria
Belgium
Brazil
Canada
Central America
Chile
China
Colombia
Czech Republic
Denmark
Egypt
Finland
France
Greece
Hong Kong
Hungary
India
Indonesia
Ireland
Israel
Italy
Jamaica
Japan
Korea
Latin America
Malaysia
Mexico
Middle East
Morocco
Netherlands
New Zealand
Norway
Pakistan
Peru
Philippines
Poland
Portugal
Puerto Rico
Russia and CIS
Saudi Arabia
Singapore
South Africa
Spain
Sweden
Switzerland
Taiwan
Thailand
Turkey
Uruguay
Venezuela
Vietnam

Autosorb[®]-1-C/MS SPECIFICATIONS

General

Range:	1 to 200 amu (optional)
Detectors:	Faraday Cup Electron Multiplier (optional)
Ion Source:	Open or Closed
Maximum Operating Pressure:	1x10 ⁻⁴ (open ion source) 1x10 ⁻² (closed ion source)
Resolution:	<10% of Valley (per AVS Standard 2.3)
Sensitivity:	2x10 ⁻⁴ amps/torr (open ion source)
Min. Detectable Partial Pressure:	Faraday Cup (2x10 ⁻¹¹ torr) Electron Multiplier (5x10 ⁻¹⁴ torr)

Physical

ECU Operating Temperature:	0°C to 40°C
-----------------------------------	-------------

Compliance

CE Mark:	EMC 89/336/EEC (electromagnetic compatibility) 72/23/EEC Low Voltage Directive
-----------------	---

SELECTED APPLICATIONS

Physisorption Studies
Micropore Analysis
Mesopore Analysis
Fractal Dimensions
Chemisorption Tests
Carbon Reactivity
Catalyst Dispersion
Surface Acidity and
Basicity
Other

CORPORATE HEADQUARTERS

Quantachrome Instruments

1900 Corporate Drive
Boynton Beach, FL 33426 USA

Phone: (561) 731-4999

Fax: (561) 732-9888

E-mail: qc.sales@quantachrome.com

EUROPE

Quantachrome UK Limited

European Technical Center
Pale Lane Farm, Pale Lane
Hartley Wintney, Hook RG27 8BA, UK

Phone: 01252 819719

Fax: 01252 819901

Quantachrome GmbH

Rudolf-Diesel Str. 12
85235 Odelzhausen, Germany

Phone: 08 134/93240

Fax: 08 134/932425



Optimizing particle performance



<http://www.quantachrome.com>

Trademarks and registered trademarks are the property of their respective owners. 07112

US Patent No. 6,257,835
Other Patents Pending
© 2001 Quantachrome Corporation

1001

Our policy of continuous development may cause the information and specifications contained herein to change without notice or liability.